



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20230706U

Expiry date: Oct. 05, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jul. 24, 2023

KBG6033100ab-Eaa
SA2307-0229

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company