



Kikkoman Biochemifa Company

**CERTIFICATE OF ANALYSIS**

**LuciPac Pen**

Product code :60331

Lot No.: 20230803U

Expiry date: Nov. 02, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	13000	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Aug. 23, 2023

KBG6033100ab-Eaa  
SA2308-0233

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company