

CERTIFICATE OF ANALYSISLuciPac Pen

Product code :60331

Lot No.: 20240131U

Expiry date: Apr. 30, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	14400	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Feb. 19, 2024KBG6033100ab-Eaa
SA2402-0184Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company