

CERTIFICATE OF ANALYSIS**LuciPac Pen**

Product code :60331

Lot No.: 20230316U

Expiry date: Jun. 15, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
-----------------------	---------------	----------------------

Average Luminescence(RLU)	13400	$\geq 9100, \leq 16900$
---------------------------	-------	-------------------------

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Ryosuke Sugimoto; ManagerDate May. 11, 2023KBG6033100ab-Eaa
SA2305-0110Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company