

CERTIFICATE OF ANALYSIS**LuciPac A3 Water**

Product code :60365

Lot No.: 20230904W

Expiry date: Dec. 03, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12000	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	11600	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Oct. 23, 2023KBG6036500ac-Eaa
SA2310-0246Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company