

**CERTIFICATE OF ANALYSIS****LuciPac A3 Water**

Product code :60365

Lot No.: 20231003W

Expiry date: Jan. 02, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11900	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; ManagerDate Nov. 22, 2023KBG6036500ac-Eaa  
SA2311-0251Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company