



CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365


Lot No.: 20240110W

Expiry date: Apr. 09, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12200	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jan. 23, 2024

KBG6036500ac-Eaa
SA2401-0304

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company