

CERTIFICATE OF ANALYSISLuciPac A3 Water

Product code :60365

Lot No.: 20240412W

Expiry date: Jul. 11, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate May. 20, 2024KBG6036500ac-Eaa
SA2405-0211Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company