

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface (Pre-Moistened)**

Product code :60367

Lot No.: 20240311M

Expiry date: Nov. 10, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	17670	$\geq 11550, \leq 21450$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	18470	$\geq 11550, \leq 21450$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Apr. 24, 2024KBG6036700aa-Eaa
SA2404-0302Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company