

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface**

Product code :60361

Lot No.: 20240305Y

Expiry date: Jun. 04, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12200	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Mar. 21, 2024KBG6036100ac-Eaa
SA2403-0229Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company