

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361

Lot No.: 20230330Y

Expiry date: Jun. 29, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate May. 11, 2023KBG6036100ac-Eaa
SA2305-0112Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company