



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20230331Y

Expiry date: Jun. 30, 2024

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|--|---------------|-------------------------|
| When applied 100 μ L of 5pmol ADP Average Luminescence(RLU) | 12700 | $\geq 9100, \leq 16900$ |
| When applied 100 μ L of 5pmol AMP Average Luminescence(RLU) | 12300 | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date May. 11, 2023

KBG6036100ac-Eaa
SA2305-0113

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company