

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361

Lot No.: 20230518Y

Expiry date: Aug. 17, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13200	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; ManagerDate Jun. 16, 2023KBG6036100ac-Eaa
SA2306-0208Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company