



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20230712Y

Expiry date: Oct. 11, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5 μ mol ADP Average Luminescence(RLU)	13600	$\geq 9100, \leq 16900$
When applied 100 μ L of 5 μ mol AMP Average Luminescence(RLU)	13400	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jul. 24, 2023

KBG6036100ac-Eaa
SA2307-0231

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company