

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361

Lot No.: 20230908Y

Expiry date: Dec. 07, 2024

| <u>TEST PARAMETER</u>  | <u>RESULT</u> | <u>SPECIFICATION</u>    |
|--|---------------|-------------------------|
| When applied 100 $\mu$ L of 5pmol ADP<br>Average Luminescence(RLU) | 12500         | $\geq 9100, \leq 16900$ |
| When applied 100 $\mu$ L of 5pmol AMP<br>Average Luminescence(RLU) | 13300         | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; ManagerDate Sep. 22, 2023KBG6036100ac-Eaa  
SA2309-0304Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company