

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361


Lot No.: 20240111Y

Expiry date: Apr. 10, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12000	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jan. 23, 2024

KBG6036100ac-Eaa
SA2401-0299

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company