

# CERTIFICATE OF ANALYSIS

## Histamine Test

Product code :61341

Lot No.: 20231207A

Expiry date: Jul. 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	100.7	100±5
Absorbance of reagent blank	0.011	≤0.04
Histamine dehydrogenase activity(U/vial)	2.22	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	101.1	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Mar. 21, 2024

KBG6044101aa-Eaa  
SA2403-0224

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company