

# CERTIFICATE OF ANALYSIS

## Easy Plate YM-R

Product code :61977

Lot No.: YM20230614A

Expiry date: Dec. 13, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

### Performance test

Aspergillus niger (NBRC 33023) : 50-500\*  
 Penicillium citrinum (NBRC 6352) : 50-500\*  
 Candida albicans (NBRC 1594) : 50-500\*  
 Saccharomyces cerevisiae (NBRC 100929) : 50-500\*  
 Bacillus subtilis (NBRC 3134) : No growth  
 Escherichia coli (NBRC 15034) : No growth  
 Pseudomonas aeruginosa (NBRC 3899) : No growth

\*The ratio of the average colony count on the product to the potato dextrose agar (PDA) medium(%).

Storage condition : 2-8°C

  
 Ryosuke Sugimoto; Manager

Date Oct. 23, 2023

KBG6197700aa-Eaa  
 SA2310-0254

Quality Assurance Group, Edogawa  
 Quality Unit  
 Kikkoman Biochemifa Company