

## CERTIFICATE OF ANALYSIS

### Easy Plate SA

Product code :61976

Lot No.: SA20231005A

Expiry date: Oct. 04, 2024

| <u>TEST PARAMETER</u>      | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test           | passed test   | See below            |
| Spreading time             | passed test   | $\leq 60$ sec        |
| Gelling of medium          | passed test   | Gelling within 3 min |
| Curl of the base sheet     | passed test   | $\leq 4$ mm          |
| Appearance of culture area | passed test   | Light yellow         |
| Scanning of QR code        | passed test   | Passed               |

#### Performance test

Staphylococcus aureus (ATCC25923) : 50-200 (as blue colony)<sup>\*1</sup>

Staphylococcus aureus (NBRC13276) : 50-200 (as blue colony)<sup>\*1</sup>

Enterococcus faecalis (NBRC100482) : No growth

Bacillus subtilis (NBRC3134) : No growth

Escherichia coli (NBRC15034) : No growth


Proteus mirabilis (NBRC105697) : No growth

Bacillus licheniformis (NBRC12200) : Growth as pink colony

Staphylococcus epidermidis (NBRC100911) :  $\leq 10$  and no growth as blue colony<sup>\*2</sup>
<sup>\*1</sup>The ratio of the average colony count on the product to the plate count agar (%).

<sup>\*2</sup>The ratio of the average colony count on the product to Easy Plate AC (%).

Storage condition : 2-8°C

  
Ryosuke Sugimoto; Manager

Date Nov. 22, 2023

KBG6197600aa-Eab  
SA2311-0262

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company