

# CERTIFICATE OF ANALYSIS

## Easy Plate EC

Product code :61975

Lot No.: EC20230123A

Expiry date: Jul. 22, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

### Performance test

Escherichia coli (NBRC15034) : 70-200\*

Klebsiella pneumoniae(NBRC14940) : 50-200\*

Enterobacter cloacae (NBRC13536) : 50-200\*

Citrobacter freundii (NBRC12681) : 50-200\*

Staphylococcus aureus (ATCC25923) : No growth

Enterococcus faecalis (NBRC100482) : No growth

Proteus mirabilis (NBRC105697) : No growth

Staphylococcus xylosus (NBRC109770) : No growth

\*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C

  
Ryosuke Sugimoto; Manager

Date Apr. 24, 2023

KBG6197500aa-Eab  
SA2304-0246

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company