

CERTIFICATE OF ANALYSIS

Easy Plate EC

Product code :61975

Lot No.: EC20230308A

Expiry date: Sep. 07, 2024

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test | passed test | See below |
| Spreading time | passed test | ≤ 60 sec |
| Gelling of medium | passed test | Gelling within 3 min |
| Curl of the base sheet | passed test | ≤ 4 mm |
| Appearance of culture area | passed test | Light yellow |
| Scanning of QR code | passed test | Passed |

Performance test

Escherichia coli (NBRC15034) : 70-200*

Klebsiella pneumoniae(NBRC14940) : 50-200*

Enterobacter cloacae (NBRC13536) : 50-200*

Citrobacter freundii (NBRC12681) : 50-200*

Staphylococcus aureus (ATCC25923) : No growth

Enterococcus faecalis (NBRC100482) : No growth

Proteus mirabilis (NBRC105697) : No growth

Staphylococcus xylosus (NBRC109770) : No growth

*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C


Ryosuke Sugimoto; Manager

Date Jun. 16, 2023

KBG6197500aa-Eab
SA2306-0201

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company