

**CERTIFICATE OF ANALYSIS****Easy Plate EB**

Product code :61978

Lot No.: EB20231225A

Expiry date: Dec. 24, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Between Orange to Yellow
Scanning of QR code	passed test	Passed

## Performance test

Escherichia coli (NBRC15034) : 50-200\*

Klebsiella pneumoniae(NBRC14940) : 50-200\*

Proteus mirabilis (NBRC105697) : 50-200\*

Salmonella enterica subsp. enterica (NBRC13245) : 50-200\*

Pseudomonas aeruginosa (NBRC3899) : No growth

Enterococcus faecalis (NBRC100482) : No growth

\*The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C

  
Ryosuke Sugimoto; ManagerDate Mar. 21, 2024KBG6197800aa-Eaa  
SA2403-0215Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company