

## CERTIFICATE OF ANALYSIS

### Easy Plate CC

Product code :61974

Lot No.: CC20231117A

Expiry date: May. 16, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

#### Performance test

Escherichia coli (NBRC15034) : 70-200<sup>\*</sup>

Klebsiella pneumoniae(NBRC14940) : 50-200<sup>\*</sup>

Enterobacter cloacae (NBRC13536) : 50-200<sup>\*</sup>

Citrobacter freundii (NBRC12681) : 50-200<sup>\*</sup>

Staphylococcus aureus (ATCC25923) : No growth

Enterococcus faecalis (NBRC100482) : No growth

Proteus mirabilis (NBRC105697) : No growth

Staphylococcus xylosus (NBRC109770) : No growth

<sup>\*</sup>The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C

  
Ryosuke Sugimoto; Manager

Date Feb. 19, 2024

KBG6197400aa-Eab  
SA2402-0183

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company