

CERTIFICATE OF ANALYSIS

Easy Plate AC

Product code :61973

Lot No.: AC20240228A

Expiry date: Aug. 27, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

*The ratio of the average colony count on the product to the plate count agar(%).
 Escherichia coli (NBRC15034) : 70-200
 Bacillus subtilis (NBRC3134) : 70-350
 Klebsiella pneumoniae (NBRC14940) : 50-200
 Pseudomonas aeruginosa (NBRC3899) : 70-350
 Staphylococcus aureus (ATCC25923) : 50-200
 Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8°C


 Ryosuke Sugimoto; Manager

Date May. 20, 2024

KBG6197300aa-Eaa
SA2405-0212

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company