

# CERTIFICATE OF ANALYSIS

## Easy Plate AC

Product code :61973

Lot No.: AC20230911A

Expiry date: Mar. 10, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

\*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 70-200  
Bacillus subtilis (NBRC3134) : 70-350  
Klebsiella pneumoniae (NBRC14940) : 50-200  
Pseudomonas aeruginosa (NBRC3899) : 70-350  
Staphylococcus aureus (ATCC25923) : 50-200  
Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8°C

  
Ryosuke Sugimoto; Manager

Date Nov. 22, 2023

KBG6197300aa-Eaa  
SA2311-0253

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company