

# Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	3G2884
Manufactured	Jun 14, 2023	doc. No.	CAMT11PBS-20230731-3
Expiration	May. 2025	doc. Issued	Jul 31, 2023

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	9.0 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.3	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD	
Authorised July 31, 2023	Inspected July 31, 2023