

# Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	3G2684
Manufactured	May 22, 2023	doc. No.	CAMT11PBS-20230731-1
Expiration	Apr. 2025	doc. Issued	Jul 31, 2023

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.8 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD	
Authorized July 31, 2023	Inspected July 31, 2023
	