

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11MRD
Volume	9ml	Lot No.	3F0177
Manufactured	Jun 15, 2023	doc. No.	CAMTT11MRD-20230731-1
Expiration	May, 2025	doc. Issued	Jul 31, 2023

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.9 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	6.8~7.2	7.1	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
July 31, 2023	July 31, 2023
	