

**ELMEX LTD**

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan  
TEL: 81-3-5229-7911 FAX: 81-3-5261-4130  
URL: <http://www.elmex.co.jp>

# Certificate of Analysis

|              |              |             |                  |
|--------------|--------------|-------------|------------------|
| Product      | Swab Test    | Code No.    | ST-25PBS         |
| Volume       | 10ml         | Lot No.     | 1A7800           |
| Manufactured | Dec 22, 2021 | Doc. No.    | CAST25PBS-220307 |
| Expiration   | Nov. 2023    | Doc. Issued | Mar 07, 2022     |

| Item               | Specification                      | Result                             | Evaluation |
|--------------------|------------------------------------|------------------------------------|------------|
| Appearance: Liquid | clear, slight yellow liquid        | clear, slight yellow liquid        | passed     |
| Appearance: Bottle | no visible particle                | no visible particle                | passed     |
| Appearance: Swab   | no visible particle                | no visible particle                | passed     |
| Labels             | correct position, correct printing | correct position, correct printing | passed     |
| Net Weight         | 10.2g±0.2g                         | 10.2 g                             | passed     |
| NaCl Content       | 0.85±0.1%                          | 0.9%                               | passed     |
| pH                 | 7.2±0.2                            | 7.4                                | passed     |
| Leakage            | no leak, -0.03Mpa hold 30min.      | no leak                            | passed     |
| Sterility Test     | no bacteria (By TGC&PDA)           | no bacteria                        | passed     |
| Remarks            |                                    |                                    |            |

ELMEX LTD

| Authorised  | Inspected   |
|---|---|
| March 7, 2022   | March 7, 2022   |
|  |  |