

**ELMEX LTD**

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan  
 TEL: 81-3-5229-7911 FAX: 81-3-5261-4130  
 URL: <http://www.elmex.co.jp>

# Certificate of Analysis

Product	Swab Test	Code No.	ST-25PBS
Volume	10ml	Lot No.	2A4200
Manufactured	Aug 08, 2022	Doc. No.	CAST25PBS-220915-2
Expiration	Jul. 2024	Doc. Issued	Sep 15, 2022

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Appearance: Swab	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	10.2g±0.2g	10.2 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.2±0.2	7.3	passed
Leakage	no leak, -0.03Mpa hold 30min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
September 15, 2022	September 15, 2022