

QC certificate  
- Certificate of Analysis -

Product name:	Procheck E-W
Lot number:	KM5E201D
Expire date:	2026/05

[1] Physical Characteristics

Appearance (cleanness)	clear
Appearance (color)	yellow

[2] QC test result

Sensitivity is good enough according to our QC standard.

Detection limit using bovine serum albumin is 0.02 mg

Turns blue or green	
0.16mg	+
0.08mg	+
0.04mg	+
0.02mg	+

Name of QC responsible person: Yukari Kataoka

Date: 21 - May - 2024

*Yukari Kataoka*

Signature of responsible person

Nikken Bio Inc.  
Teigaiberi 23 Oohashiberi Kumiya-ma-cho Kuze-gun KYOTO 613-0046 JAPAN

