



CERTIFICATE OF ANALYSIS

Histamine Check Swab


Product code :60448

Lot No.: 20230327J

Expiry date: Sep. 26, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Absorbance(600nm) of 100ppm histamine solution	1.66	1.55-1.95

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date Apr. 24, 2023

KBG6044800aa-Eaa
SA2304-0245

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company