

CERTIFICATE OF ANALYSIS**Histamine Check Swab**

Product code :60448

Lot No.: 20231229J

Expiry date: Jun. 28, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Absorbance(600nm) of 100ppm histamine solution	1.64	1.55-1.95

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jan. 23, 2024KBG6044800aa-Eaa
SA2401-0305Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company