



CERTIFICATION

AOAC Research Institute *Performance Tested Methods*SM

Certificate No.
041302

The AOAC Research Institute hereby certifies the method known as:

Easy Plate AC

manufactured by

Kikkoman Biochemifa Company
2-1-1, Nishi-shinbashi
Minato-ku, Tokyo 105-0003
Japan

This method has been evaluated in the AOAC Research Institute *Performance Tested Methods*SM Program and found to perform as stated in the applicability of the method. This certificate indicates an AOAC Research Institute Certification Mark License Agreement has been executed which authorizes the manufacturer to display the AOAC Research Institute *Performance Tested Methods*SM certification mark on the above-mentioned method for the period below. Renewal may be granted by the Expiration Date under the rules stated in the licensing agreement.

A handwritten signature in black ink that reads "Scott Coates".

Scott Coates, Senior Director
Signature for AOAC Research Institute

Issue Date	November 28, 2022
Expiration Date	December 31, 2023

2275 Research Blvd., Suite 300, Rockville, MD 20850-3250 USA * Telephone: +1-301-924-7077 * Fax: +1-301-924-7089

Internet e-mail: aoacri@aoac.org * World Wide Web Site: <http://www.aoac.org>

AUTHORS ORIGINAL VALIDATION: Norihiko Okochi, Mamoru Yamazaki, Shoichi Kiso, Mai Kinoshita, Yurie Okita, Keisuke Kazama, and Rui Saito MODIFICATION DECEMBER 2017: Mai Shimizu, Kentaro Takenaka, Aya Miyasaka, Takeo Suzuki, Taiki Matsuda, Tatsuhiko Iwase, and Hitoshi Kyotani MODIFICATION MARCH 2022: Satoshi Fukuda and Shinichiro Sugiura		SUBMITTING COMPANY Dai Nippon Printing Co., Ltd. 1-1-1, Ichigaya Kagacho Shinjuku-ku, Tokyo, 162-8001 Japan	CURRENT COMPANY Kikkoman Biochemifa Company 2-1-1, Nishi-shinbashi Minato-ku, Tokyo 105-0003 Japan
METHOD NAME Easy Plate AC Formerly known as Medi-Ca AC		CATALOG NUMBER 61973	
INDEPENDENT LABORATORY Q Laboratories, Inc 1400 Harrison Avenue Cincinnati, OH 45214 USA		AOAC EXPERTS AND PEER REVIEWERS Yi Chen ¹ , Yvonne Salfinger ² , Wayne Ziemer ³ ¹ Food and Drug Administration, Center for Food Safety and Applied Nutrition, Maryland, USA ² Consultant, Colorado, USA ³ Consultant, Georgia, USA Modifications February 2020 and March 2022 reviewed internally by AOAC Research Institute.	
APPLICABILITY OF METHOD Target organism – Aerobic bacteria Matrixes – (50) – cooked roast beef, Chinese barbecued pork, bacon, cooked ham, frankfurter (made from beef and pork), boiled and cooked pork sausage MODIFICATION DECEMBER 2017: (50 g): raw beef (steak meat), raw ground beef (80 % lean), raw ground pork, raw chicken, raw ground chicken, bagged lettuce, blueberries, raw shrimp (without head), raw salmon filet, raw tuna filet, dry pet food pasteurized milk (3.6 % fat, 11 mL), natural cheese (Camembert, 11 g) Performance claims – The method is a reasonable alternative to the Microbiological Methods (AOAC Official Method 966.23) (3) for heat-processed meat matrixes.		REFERENCE METHODS <i>Official Methods of Analysis</i> (2012) 19 th Ed., AOAC INTERNATIONAL, Gaithersburg, MD, Method 966.23 (3) U.S. Food and Drug Administration (2013) <i>FDA Bacteriological Analytical Manual</i> , Chapter 3 (9)	
ORIGINAL CERTIFICATION DATE April 10, 2013		CERTIFICATION RENEWAL RECORD Renewed annual through December 2023.	
METHOD MODIFICATION RECORD <ol style="list-style-type: none"> January 2016 Level 2 December 2017 Level 3 February 2020 Level 2 November 2020 Level 1 June 2021 Level 1 March 2022 Level 2 		SUMMARY OF MODIFICATION <ol style="list-style-type: none"> Shelf life extension approved . Matrix extension Manufacturing location change from Tokyo, Japan to Kanagawa, Japan. Editorial and formatting changes to insert. Rebranded kit to reflect Kikkoman and method name change from Medi-Ca AC to Easy Plate AC. Manufacturing location change. 	
Under this AOAC <i>Performance Tested Methods</i> SM License Number, 041302 this method is distributed by: <ol style="list-style-type: none"> AS ONE Corporation KENIS LIMITED Nippon Bacterial Test CO., LTD. FUJIFILM Wako Pure Chemical Corporation Microgiene Co. LTD Weber Scientific ELMEX Limited 		Under this AOAC <i>Performance Tested Methods</i> SM License Number, 041302 this method is distributed as: <ol style="list-style-type: none"> Easy Plate AC Easy Plate AC Easy Plate AC Easy Plate AC Easy Plate AC Easy Plate AC Easy Plate AC 	

PRINCIPLE OF THE METHOD (1)

Easy Plate AC (formerly known as Medi-Ca AC) is a ready-made dry medium for aerobic count made up of four components: a waterproof sheet, a dry medium containing a gelling agent, a hydrophobic resin ring surrounding the medium, and a transparent cover over the medium. The cover is lifted, sample suspension is placed on the center of the medium, and the cover is dropped onto the sample. The sample soaks into the medium and turns into a gel in 3 minutes. The gelled medium contains the redox indicator 2,3,5-triphenyl tetrazolium chloride (TTC) derived from a coating on the cover. The incubation of the sheet at $35 \pm 1^\circ\text{C}$ for 48 ± 2 hours develops red colonies because of the redox reaction involving the indicator.

DISCUSSION OF THE VALIDATION STUDY (1)

The Easy Plate AC method was compared to the AOAC **966.23** method for seven different heat-processed meat matrixes in the two Matrix studies. The 95% CI for the mean difference between the two methods at each contamination level for each matrix fell within the range of -0.50 to 0.50, and no statistical difference was observed at all three contamination levels for five matrixes (Table 1 and 4). In addition, the repeatability of the Easy Plate AC method was overall similar to that of the AOAC **966.23** method. These results demonstrated that the Easy Plate AC method is a reasonable alternative to the AOAC **966.23** method for heat-processed meat matrixes.

The mean \log_{10} counts of the Medi-Ca AC method for boiled pork sausage contaminated with the heat-stressed strain were significantly lower than those of the AOAC **966.23** method. Interestingly, the same matrix contaminated with the same strain without any heat treatment provided no significant difference (data not shown).

These results suggest that the Medi-Ca AC medium cannot grow heat-stressed microorganisms as vigorously as the PCA, depending on the microorganisms.

The Medi-Ca AC method is similar to the Aerobic Plate Count in Foods (AOAC Official Method **990.12**), also known as the Petrifilm™ Aerobic Count Plate method (4). Morita et al. has pointed out that liquefaction of the gel by bacteria which caused diffusion of colonies was observed on the Petrifilm™ Aerobic Count plates (5). The same phenomenon, which sometimes interfered with counting, was also observed for Chinese barbecued pork, cooked ham, and frankfurter in this study (data not shown). On the other hand, Medi-Ca AC medium did not appear to be subject to the liquefaction by bacteria for all the matrixes, which made counting easier.

Table 1. Matrix Study (Method Developer) (1) *Medi-Ca AC is now Easy Plate AC

Matrix	Inoculation Micoorganism	Contamination Level	Medi-Ca AC			AOAC 966.23			Mean Difference	95% CI ^d		r ^{2g}
			Mean ^a	s _r ^b	RSD _r ^c	Mean	s _r	RSD _r		LCL ^e	UCL ^f	
Cooked roast beef	N/A ^h	Low	3.51	0.03	0.89	3.44	0.10	2.97	-0.07	-0.23	0.10	0.99
		Medium	6.20	0.14	2.31	6.28	0.08	1.26	0.08	-0.04	0.19	
		High	8.59	0.15	1.69	8.63	0.14	1.67	0.04	-0.31	0.39	
Chinese barbecued pork	N/A	Low	4.61	0.07	1.58	4.56	0.03	0.61	-0.05	-0.13	0.04	1.00
		Medium	7.93	0.05	0.64	8.00	0.04	0.52	0.07	0.00	0.14	
		High	8.56	0.06	0.65	8.64	0.06	0.64	0.08	0.00	0.16	
Bacon	N/A	Low	4.34	0.03	0.66	4.34	0.05	1.13	0.01	-0.05	0.07	0.99
		Medium	6.35	0.02	0.32	6.30	0.04	0.60	-0.04	-0.11	0.03	
		High	7.43	0.08	1.03	7.57	0.06	0.74	0.14	-0.01	0.28	
Cooked ham	N/A	Low	2.61	0.04	1.57	2.60	0.04	1.40	-0.01	-0.10	0.08	0.99
		Medium	7.09 ⁱ	0.04	0.54	7.48	0.04	0.54	0.39	0.34	0.43	
		High	9.26 ⁱ	0.03	0.35	9.12	0.06	0.70	-0.14	-0.25	-0.03	
Frankfurter	N/A	Low	4.88	0.05	0.93	4.91	0.04	0.78	0.04	-0.04	0.11	0.99
		Medium	5.74	0.04	0.70	5.71	0.05	0.82	-0.03	-0.07	0.01	
		High	6.12	0.04	0.59	6.13	0.03	0.55	0.02	-0.04	0.08	
Boiled pork sausage	<i>E. cloacae</i> ^j ATCC 222	Uninoculated	<1.00	—	—	<1.00	—	—	—	—	—	0.99
		Low	2.60 ⁱ	0.06	2.18	2.97	0.04	1.47	0.37	0.27	0.47	
		Medium	3.58 ⁱ	0.04	1.11	3.81	0.05	1.40	0.24	0.12	0.35	
		High	4.55 ⁱ	0.09	1.87	4.74	0.05	0.96	0.19	0.05	0.33	

^a Mean of 5 replicates after the logarithmic transformation: Log₁₀[CFU/g + (0.1)^f].^b s_r = standard deviation.^c RSD_r = relative standard deviation.^d CI = confidence interval.^e LCL = lower confidence limit.^f UCL = upper confidence limit.^g r² = square of the correlation coefficient.^h N/A—Not applicable. Samples are naturally contaminated.ⁱ Significantly different (p<0.05).^j A heat-stressed culture with 71% injury was used.

DISCUSSION OF A MODIFICATION STUDY APPROVED DECEMBER 2017 (8)

Easy Plate AC method was compared to the AOAC 966.23 and BAM Chapter 3 for 13 food matrixes in five different categories: meat products, vegetable and fruits, seafood, dairy products and pet food including heat processed food. According to the validation results in method developer study and independent laboratory study, the mean differences between the Easy Plate AC and reference methods were less than 0.16 log₁₀, and much smaller in most cases. The 95% CIs for the mean differences between the two methods fell within the range of -0.5 to 0.5 (Table 1). These results demonstrated that the Medi-Ca AC method produced statistically similar results when compared to the reference method.

Furthermore, for the dairy products (pasteurized milk and natural cheese), incubation temperature in the reference method is 32°C. In this study, Medi-Ca AC was incubated at 32 and 35 ± 1°C in the Method Developer Study for the dairy products and only the pasteurized milk in the Independent Study. As a result, there were no differences in these two incubation temperature (Table 1). Therefore, both dairy and non-dairy foods can be incubated in 35 ± 1°C by using Medi-Ca AC, and it is not necessary to use a different incubator for each food type.

Table 1. Matrix Study: Medi-Ca AC vs. AOAC 966.23 and BAM Chapter 3 (1) *Medi-Ca AC is now Easy Plate AC

Matrix	Inoculation Microorganism	Contamination Level	Medi-Ca AC			AOAC 966.23			Mean Difference	95% CI ^d		r ^{2g}
			Mean ^a	s _r ^b	RSD _r ^c	Mean	s _r	RSD _r		LCL ^e	UCL ^f	
Raw beef	N/A ^h	Low	4.93	0.03	0.68	4.91	0.08	1.53	0.02	-0.12	0.08	0.998
		Medium	6.37	0.03	0.44	6.35	0.03	0.49	0.02	-0.04	0.01	
		High	7.15	0.02	0.30	7.12	0.04	0.57	0.03	-0.06	0.01	
Raw beef ⁱ	N/A	Low	1.88	0.19	10.1	1.92	0.17	8.83	0.04	-0.24	0.15	0.979
		Medium	2.95	0.15	5.10	3.02	0.11	3.72	0.08	-0.18	0.03	
		High	3.76	0.12	3.21	3.75	0.14	3.65	0.01	-0.14	0.15	
Raw ground beef	N/A	Low	6.90	0.03	0.50	6.85	0.05	0.75	0.05	-0.11	0.02	0.996
		Medium	7.87	0.05	0.67	7.81	0.04	0.50	0.06	-0.15	0.04	
		High	8.81	0.04	0.44	8.79	0.04	0.48	0.02	-0.06	0.02	
Raw ground pork	N/A	Low	5.47	0.05	0.94	5.63	0.11	1.91	0.16	0.02	0.29	0.998
		Medium	8.93	0.06	0.67	8.95	0.06	0.65	0.03	-0.10	0.15	
		High	9.77	0.01	0.14	9.80	0.04	0.41	0.03	-0.02	0.08	
Raw chicken	N/A	Low	4.33	0.04	0.97	4.41	0.08	1.73	0.08	-0.05	0.20	0.998
		Medium	8.73	0.04	0.42	8.63	0.09	1.00	0.11	-0.22	0.01	
		High	9.63	0.07	0.70	9.60	0.04	0.45	0.03	-0.13	0.07	
Raw ground chicken	N/A	Low	7.34	0.04	0.49	7.33	0.03	0.47	0.00	-0.09	0.08	0.996
		Medium	8.35	0.02	0.28	8.34	0.02	0.21	0.01	-0.06	0.04	
		High	9.37	0.03	0.28	9.29	0.03	0.27	0.08	-0.13	-0.03	
Lettuce	N/A	Low	4.63	0.08	1.70	4.46	0.05	1.17	0.16	-0.22	-0.11	0.997
		Medium	6.02	0.03	0.44	6.02	0.06	0.97	0.00	-0.07	0.07	
		High	7.74	0.08	1.07	7.74	0.04	0.58	0.01	-0.11	0.12	
Blueberries	N/A	Low	2.61	0.05	1.90	2.76	0.06	2.24	0.15	0.03	0.26	0.987
		Medium	3.26	0.02	0.70	3.29	0.06	1.90	0.02	-0.05	0.10	
		High	4.47	0.07	1.60	4.39	0.05	1.18	0.08	-0.21	0.05	
Raw shrimp	N/A	Low	7.30	0.02	0.29	7.29	0.04	0.50	0.01	-0.05	0.03	0.995
		Medium	8.15	0.03	0.36	8.19	0.04	0.48	0.04	-0.04	0.11	
		High	9.19	0.05	0.59	9.25	0.03	0.34	0.06	-0.03	0.15	
Raw salmon	N/A	Low	7.57	0.09	1.15	7.66	0.05	0.70	0.09	-0.03	0.20	0.990
		Medium	8.65	0.06	0.65	8.68	0.07	0.77	0.03	-0.07	0.12	
		High	9.63	0.07	0.76	9.62	0.05	0.50	0.01	-0.15	0.13	
Raw tuna	N/A	Low	6.96	0.03	0.40	6.90	0.03	0.44	0.06	-0.09	-0.03	0.997
		Medium	7.89	0.04	0.53	7.90	0.03	0.41	0.01	-0.03	0.06	
		High	8.84	0.05	0.55	8.87	0.03	0.33	0.03	-0.05	0.10	
Dry pet food ⁱ	N/A	Low	1.76	0.15	8.38	1.76	0.16	9.18	0.00	-0.12	0.13	0.996
		Medium	3.98	0.07	1.64	3.96	0.07	1.86	0.02	-0.08	0.13	
		High	4.88	0.07	1.42	4.96	0.02	0.44	0.08	-0.14	-0.01	

Pasteurized milk 32°C ^j	<i>Escherichia coli</i> NBRC ^k 13500	Uninoculated	<1.00			<1.00						0.977
		Low	2.52	0.08	3.33	2.57	0.04	1.70	0.04	-0.01	0.10	
		Medium	3.62	0.07	2.00	3.50	0.08	2.17	0.12	-0.21	-0.03	
		High	4.57	0.12	2.58	4.52	0.09	2.07	0.05	-0.30	0.19	
Pasteurized milk 35°C ^j	<i>E. coli</i> NBRC 13500	Uninoculated	<1.00			<1.00						0.988
		Low	2.55	0.07	2.81	2.57	0.04	1.70	0.02	-0.04	0.08	
		Medium	3.58	0.07	2.09	3.50	0.08	2.17	0.08	-0.26	0.10	
		High	4.58	0.05	1.10	4.52	0.09	2.07	0.06	-0.14	0.02	
Pasteurized milk ⁱ 32°C ^j	N/A	Low	1.36	0.22	15.9	1.39	0.09	6.67	0.03	-0.33	0.27	0.991
		Medium	3.91	0.03	0.71	3.92	0.05	1.30	0.01	-0.09	0.06	
		High	4.92	0.02	0.47	4.85	0.08	1.70	0.07	-0.03	0.16	
Pasteurized milk ⁱ 35°C ^j	N/A	Low	1.41	0.13	9.41	1.39	0.09	6.67	0.02	-0.04	0.09	0.999
		Medium	3.90	0.06	1.61	3.92	0.05	1.30	0.02	-0.10	0.06	
		High	4.90	0.05	0.97	4.85	0.08	1.70	0.05	-0.02	0.11	
Natural cheese 32°C ^j	<i>Staphylococcus aureus</i> ATCC ^k 12600	Uninoculated	<1.00			<1.00						0.985
		Low	2.86	0.04	1.39	2.78	0.07	2.70	0.08	-0.21	0.04	
		Medium	4.01	0.07	1.84	3.91	0.03	0.75	0.10	-0.21	0.01	
		High	4.70	0.08	1.66	4.67	0.10	2.11	0.03	-0.17	0.11	
Natural cheese 35°C ^j	<i>S. aureus</i> ATCC 12600	Uninoculated	<1.00			<1.00						0.993
		Low	2.87	0.04	1.41	2.78	0.07	2.70	0.09	-0.15	-0.02	
		Medium	4.01	0.05	1.16	3.91	0.03	0.75	0.10	-0.18	-0.02	
		High	4.74	0.02	0.48	4.67	0.10	2.11	0.07	-0.18	0.04	

^a Mean of 5 replicates after the logarithmic transformation: Log₁₀[CFU/g + (0.1)^f].

^b *s_r* = standard deviation.

^c RSD_r = relative standard deviation.

^d CI = confidence interval.

^e LCL = lower confidence limit.

^f UCL = upper confidence limit.

^g *r*² = square of the correlation coefficient.

^h N/A = Not applicable. Samples are naturally contaminated.

ⁱ Matrix study conducted by the independent laboratory.

^j Incubation temperature for Medi-Ca AC

^k Biological Resource Center, National Institute of Technology and Evaluation, Chiba, Japan

^l American Type Culture Collection, Manassas, VA

REFERENCES CITED

1. Okochi, Norihiko, Yamazaki, Mamoru, Kiso, Shoichi, Kinoshita, Mai, Okita, Yurie, Kazama, Keisuke, and Saito, Rui., Evaluation of the Dai Nippon Printing Co., LTD. Medi-Ca AC for the Detection of Aerobic bacteria, AOAC Performance Tested MethodsSM certification number 041302.
2. AOAC Research Institute Validation Outline for Dai Nippon Printing Co., LTD. Medi-Ca AC, Approved – April 2013.
3. *Official Methods of Analysis* (2012) 19th Ed., AOAC INTERNATIONAL, Gaithersburg, MD, Method **966.23**
4. U.S. Food and Drug Administration (2013) *Bacteriological Analytical Manual Online*, <http://www.fda.gov/Food/FoodScienceResearch/LaboratoryMethods/ucm2006949.htm>.
5. AOAC INTERNATIONAL Methods Committee Guidelines for Validation of Qualitative and Quantitative Food Microbiological Official Methods of Analysis, AOAC INTERNATIONAL. (2011) Pre-publication.
6. *Official Methods of Analysis* (2012) 19th Ed., AOAC INTERNATIONAL, Gaithersburg, MD, Method **990.12**
7. Morita, H., Ushiyama, M., Aoyama, S., & Iwasaki, M. (2003) *J. AOAC Int.* **86**, 355-66
8. Shimizu, M., Takenaka, K., Miyasaka, A., Suzuki, T., Matsuda, T., Iwase, T., and Kyotani, H., Evaluation of Medi-Ca AC Method Modification (Matrix Extension), AOAC® Performance TestedSM certification number 041302. Approved December 2017
9. U.S. Food and Drug Administration (2013) *FDA Bacteriological Analytical Manual*, Chapter 3, (<https://www.fda.gov/Food/FoodScienceResearch/LaboratoryMethods/ucm063346.htm>), Accessed August 2017
10. Fukuda, S. and Sugiura, S., Modification Validation of Easy Plate Series (AC, CC, EC, and SA) for the Manufacturing Location Change, AOAC Performance Tested MethodsSM certification number 041302. Approved March 2022.