

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484
Product name: **Lumitester Calibration Kit**
Serial Number: 23K043

DATA

for PD-30/PD-20

RLU value on the label: 1,380 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,370 RLU
Difference: 0.7 %
(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,740 RLU
RLU value obtained by a reference Lumitester Smart: 1,720 RLU
Difference: 1.1 %
(Std.: within ± 10 %)

Date: Jul.23,2024Signature: H Katayama
Hiroshi Katayama; ManagerQuality Assurance Group, Edogawa
Quality Unit**Kikkoman Biochemifa Company**

KBG60484-Eac

SA2407047