

**CERTIFICATE OF ANALYSIS**

Product code (Kikkoman): 60484  
Product name: **Lumitester Calibration Kit**  
Serial Number: 23K042

**DATA**

for PD-30/PD-20

RLU value on the label: 1,070 RLU  
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,070 RLU  
Difference: 0.0 %  
(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,320 RLU

RLU value obtained by a reference Lumitester Smart: 1,330 RLU  
Difference: -0.8 %  
(Std.: within ± 10 %)

Date: Jul.23,2024

Signature: H Katayama  
Hiroshi Katayama; Manager

Quality Assurance Group, Edogawa  
Quality Unit

**Kikkoman Biochemifa Company**

KBG60484-Eac  
SA2407046