

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484
Product name: **Lumitester Calibration Kit**
Serial Number: 23K041

DATA

for PD-30/PD-20

RLU value on the label: 930 RLU
(Std.: within 1100 \pm 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 920 RLU
Difference: 1.1 %
(Std.: within \pm 10 %)

for Smart

RLU value on the label: 1,110 RLU
RLU value obtained by a reference Lumitester Smart: 1,120 RLU
Difference: -0.9 %
(Std.: within \pm 10 %)

Date: Jul.23,2024

Signature: H Katayama
Hiroshi Katayama; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2407045