

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484
Product name: Lumitester Calibration Kit
Serial Number: 23K040

DATA

for PD-30/PD-20

RLU value on the label: 970 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 980 RLU
Difference: -1.0 %
(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,170 RLU
RLU value obtained by a reference Lumitester Smart: 1,190 RLU
Difference: -1.7 %
(Std.: within ± 10 %)

Date: Jul.23,2024Signature: H Katayama
Hiroshi Katayama; ManagerQuality Assurance Group, Edogawa
Quality Unit**Kikkoman Biochemifa Company**

KBG60484-Eac

SA2407044