

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484
Product name: **Lumitester Calibration Kit**
Serial Number: 23K037

DATA

for PD-30/PD-20

RLU value on the label: 1,260 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,270 RLU
Difference: -0.8 %
(Std.: within ±10 %)

for Smart

RLU value on the label: 1,570 RLU
RLU value obtained by a reference Lumitester Smart: 1,580 RLU
Difference: -0.6 %
(Std.: within ±10 %)

Date: Jul.23,2024

Signature: H Katayama
Hiroshi Katayama; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2407041