



## CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484  
Product name: Lumitester Calibration Kit  
Serial Number: 23K034

### DATA

for PD-30/PD-20

RLU value on the label: 1,110 RLU  
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,120 RLU  
Difference: -0.9 %  
(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,370 RLU  
RLU value obtained by a reference Lumitester Smart: 1,390 RLU  
Difference: -1.5 %  
(Std.: within ± 10 %)

Date: Jul.23,2024

Signature: H Katayama  
Hiroshi Katayama; Manager

Quality Assurance Group, Edogawa  
Quality Unit

**Kikkoman Biochemifa Company**

KBG60484-Eac

SA2407038