

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code: 60361

Lot No.: 20240703Y		Expiry date: Oct. 02, 2025	
TEST PARAMETER	RESULT	SPECIFICATION	
When applied 100µL of 5pmol ADP Average Luminescence(RLU)	11900	≥9100, ≤16900	
When applied 100µL of 5pmol AMP Average Luminescence(RLU)	13100	≥9100, ≤16900	

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Ryosuke Sugimoto; Manager

Date Jul. 23, 2024

KBG6036100ac-Eaa SA2407-0240 Quality Assurance Group, Edogawa Quality Unit Kikkoman Biochemifa Company