

CERTIFICATE OF ANALYSIS

Easy Plate EB

Product code :61978

Lot No.: EB20240612A

Expiry date: Jun. 11, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Between Orange to Yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 50-200*

Klebsiella pneumoniae(NBRC14940) : 50-200*

Proteus mirabilis (NBRC105697) : 50-200*

Salmonella enterica subsp. enterica (NBRC13245) : 50-200*

Pseudomonas aeruginosa (NBRC3899) : No growth

Enterococcus faecalis (NBRC100482) : No growth

*The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C

H. Katayama
Hiroshi Katayama, Manager

Date Sep. 19, 2024

KBG6197800aa-Eaa
SA2409-0242

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company