

CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

| PRODUCT | LOT NUMBER | EXPIRY DATE | CERTIFICATION DATE |
|-----------|------------|-------------|--------------------|
| MW503 125 | 075112 | 30/04/2025 | 24/04/2024 |

Coliform Isolation Transwab®

| PARAMETER | SPECIFICATION | RESULT |
|----------------------|--|--------|
| Sterilisation Method | Irradiation | PASSED |
| Sterility Test | Visual inspection following incubation | PASSED |
| Performance Testing | <p>Colour change induced by growth of following organisms:</p> <p><i>Escherichia coli</i> NCTC 9001 (24hr)</p> <p><i>Proteus mirabilis</i> NCTC 11938 (24hr)</p> <p>No colour change induced by growth of following organisms</p> <p><i>Staphylococcus aureus</i> NCTC 8532 (24hr)</p> <p><i>Citrobacter freundii</i> NCTC 9750 (24hr)</p> | PASSED |

Product has been manufactured in accordance with ISO 9001:2015

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance