

CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

| PRODUCT | LOT NUMBER | EXPIRY DATE | CERTIFICATION DATE |
|------------|------------|-------------|--------------------|
| MW572M 125 | 070199 | 31/03/2025 | 05/04/2024 |

Salmonella Isolation Transwab®

| PARAMETER | SPECIFICATION | RESULT |
|----------------------|--|--------|
| Sterilisation Method | Irradiation | PASSED |
| Sterility Test | Visual inspection following incubation | PASSED |
| Performance Testing | Colour change induced by growth of following organisms: <i>Salmonella enteritidis subsp enterica</i> NCTC 4444 <i>Salmonella typhimurium</i> NCTC 0074 No colour change induced by growth of following organisms <i>Escherichia coli</i> NCTC 9001 <i>Staphylococcus aureus</i> NCTC 8532 (24hr) <i>Enterococcus faecalis</i> ATCC® 19433 (24hr) | PASSED |

Product has been manufactured in accordance with ISO 9001:2015

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance