

CERTIFICATE OF ANALYSIS**Histamine Test**

Product code :61341


Lot No.: 20240325A

Expiry date: Oct. 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	100.1	100±5
Absorbance of reagent blank	0.011	≤0.04
Histamine dehydrogenase activity(U/vial)	2.16	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	99.4	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jul. 23, 2024KBG6044101aa-Eaa
SA2407-0243Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company