

Certificate of Analysis

Product	Swab Test	Code No.	ST-25PBS
Volume	10ml	Lot No.	2A0100K
Manufactured	Feb 10, 2022	Doc. No.	CAST25PBS-220602-1
Expiration	Jan. 2024 ext May 2025	Doc. Issued	Jun, 2022; Jun2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Appearance: Swab	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	10.2g±0.2g	10.3 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.2±0.2	7.1	passed
Leakage	no leak, -0.03Mpa hold 30min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
June 2, 2022	June 2, 2022
	