

EU DECLARATION OF CONFORMITY

Manufacturer: Kikkoman Biochemifa Company
2-1-1, Nishi-Shinbashi,
Minato-ku,
Tokyo, Japan

Product: Lumitester Smart

This declaration of conformity is issued under the sole responsibility of the manufacturer.


The product above is in conformity with the Union harmonization legislation:

Directive 2014/53/EU
Directive 2011/65/EU

The relevant harmonized standards used to declare the conformity with Directive 2014/53/EU are:

Safety	Art. 3 (1) a EN 62368-1:2014
Health	Art. 3 (1) a EN 62479:2010
EMC	Art. 3 (1) b EN 301489-1 V2.1.1 (2017-02)
EMC	Art. 3 (1) b EN 55032:2015
EMC	Art. 3 (1) b EN 61326-1:2013
Radio	Art. 3 (2) EN 300 328 V2.2.2

Tokyo, July 1st, 2022



Hiroshi Katayama
Manager Quality Unit
Kikkoman Biochemifa
Company

EU KONFORMITÄTSERKLÄRUNG

Hersteller: Kikkoman Biochemifa Company
2-1-1, Nishi-Shinbashi,
Minato-ku,
Tokyo, Japan

Produkt: Lumitester Smart

Die alleinige Verantwortung für diese EU Konformitätserklärung trägt der Hersteller.


Das Produkt erfüllt die einschlägigen Harmonisierungsrechtsvorschriften der Union:

Richtlinie 2014/53/EU
Richtlinie 2011/65/EU

Folgende harmonisierten Normen wurden zugrundegelegt für die Erklärung der Konformität mit der Richtlinie 2014/53/EU:

Safety	Art. 3 (1) a	EN 62368-1:2014
Health	Art. 3 (1) a	EN 62479:2010
EMC	Art. 3 (1) b	EN 301489-1 V2.1.1 (2017-02)
EMC	Art. 3 (1) b	EN 55032:2015
EMC	Art. 3 (1) b	EN 61326-1:2013
Radio	Art. 3 (2)	EN 300 328 V2.2.2

Tokyo, July 1st, 2022



Hiroshi Katayama
Manager Quality Unit
Kikkoman Biochemifa Company

Information according to Art. 8 Directive 2014/53/EU:


The product Lumitester Smart operates in the frequency band

2.4 – 2.4835GHz

with a maximum radio-frequency power of

0.14mW EIRP

Tokyo, July 1st, 2022



Hiroshi Katayama
Manager Quality Unit
Kikkoman Biochemifa Company

Information nach Art. 8 Richtlinie 2014/53/EU:


Das Produkt Lumitester Smart arbeitet in dem Frequenzband

2.4 – 2.4835GHz

mit einer maximalen Sendeleistung von

0.14mW EIRP

Tokyo, July 1st, 2022



Hiroshi Katayama
Manager Quality Unit
Kikkoman Biochemifa Company