

# CERTIFICATE OF ANALYSIS

## LuciPac A3 Surface

Product code :60361

Lot No.: 20240801Y

Expiry date: Oct. 31, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11800	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12400	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama  
Hiroshi Katayama (Manager)

Date Aug. 23, 2024

KBG6036100ac-Eaa  
SA2408-0228

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company