



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen-AQUA**

Product code :60336

---

Lot No.: 20240426Q

Expiry date: Jul. 25, 2025


---

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; Manager

Date Jun. 24, 2024

KBG6033600ab-Eaa  
SA2406-0211

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company