

## Certificate of Analysis

Certificate Number 8375

REF: ISO 15189: 2012

**Product**  
**ID-No.**  
**Lot Number**  
**Expiry Date**

**Listeria Swab**  
**1 000 100**  
**083127**  
**30.06.2025**

Test	Specifications	Results*
Appearance	Light brown colour	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
<b>Colony growth</b>		
<i>Listeria monocytogenes</i> NCTC 5214	Black colour development	Corresponds
<i>Listeria innocua</i> NCTC 11288	Black colour development	Corresponds
<i>Listeria monocytogenes</i> NCTC 11994	Black colour development	Corresponds
<i>Listeria monocytogenes</i> 7973	Black colour development	Corresponds
<i>Staphylococcus aureus</i> NCTC 8532	No colour development	Corresponds
<i>Staphylococcus faealis</i> NCTC 0775	No colour development	Corresponds
<i>Lactobacillus spec</i>	No colour development	Corresponds
<i>Listeria grayli</i> NCTC 10815	No colour development	Corresponds

**This lot meets the specifications.**

Date: 04.04.2023

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.